



Republic of the Philippines
SUGAR REGULATORY ADMINISTRATION
Araneta St., Singcang, Bacolod City 6100 Philippines

March 5, 2025

Date

REQUEST FOR QUOTATION

Please quote your LOWEST NET PRICE to the Government on the items listed below, giving full and detailed description of your offers, stating the shortest time of delivery and submit your quotation in a sealed envelope, addressed to the Sugar Regulatory Administration.

Deadline of Submission: **Not later than**

24 MAR 2025

Very truly yours,

Note: Return this form duly accomplished in a sealed envelope.

For canvassing only.

EVELYN D. PORTALLO

Administrative Officer III/RBAC Secretariat

PR No.: 2025-02-0250			DESCRIPTION	Offer
No.	Unit	Qty.		
	unit	1	<u>Research Development and Extension</u> <u>ARD - SOILS LAB BACOLOD</u> Analytical Balance (see attached specifications)	
ABC: Php: 300,000.00				
Performance Period				
Payment Term:				
Charge to 25-07583-02				

Please indicate:

In case of purchase of goods: TIN _____ Place of Delivery _____ Date of Delivery _____

Delivery Term _____ Payment Term _____

In case of labor: TIN _____ Completion Period _____ Warranty Period _____

SUGAR REGULATORY ADMINISTRATION

This is to certify that the price/s quoted above is the lowest we can offer, taxes included and that no monetary consideration, gift or in hand whatever is involved in case part or whole items quoted will be awarded to us.

The prices quoted is good until _____ and can be delivered within _____ working calendar days from receipt of the Purchase Order.

Canvassed by:

Date

(Owner/Authorized Rep.)

Telephone No.

Date

SRA - Visayas / Prop. & Procurement File / 2019

Series: _____

ANALYTICAL BALANCE

Budget: 300,000

Limit Values	
Capacity	152 g
Readability	0.1 mg
Capacity of fine range	62 g
Readability in fine range	0.01 mg
Repeatability (at 5% load)	0.03 mg
Linearity deviation	0.2 mg
Sensitivity offset (at nominal load)	0.8 mg
Typical values	
Repeatability (at 5% load)	0.015 mg
Linearity deviation	0.06 mg
Sensitivity offset (at nominal load)	0.5 mg
Minimum weight (USP, tolerance=0.10%)	30 mg
Minimum weight (tolerance=1%)	3 mg
Setting time	4s
Dimensions and Other specifications	
Balance dimensions (WxDxH)	209 x 354 x 354 mm
Weighing pan diameter)	80 mm