



Republic of the Philippines
Department of Agriculture

SUGAR REGULATORY ADMINISTRATION

Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: srahead@sra.gov.ph

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376

SUGAR REGULATORY ADMINISTRATION

North Avenue., Diliman, Quezon City

BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL BID BULLETIN NO. 1

This **Supplemental Bid Bulletin No. 1** dated **17 February 2025** is issued to inform the prospective bidders of the following amendments, modifications and clarifications on the items in the Philippine Bidding Documents (PBDs) for the **SUPPLY AND DELIVERY OF ONE (1) UNIT CARBON-NITROGEN-SULFUR (CNS) ANALYZER (RE-BIDDING)** as previously posted in the Philippine Government Electronic Procurement System (PhilGEPS) and the Sugar Regulatory Administration (SRA) website. This Supplemental Bid Bulletin shall form an integral part of the Bidding Documents.

ORIGINAL BID DOCUMENT	CLARIFICATIONS/CORRECTIONS/CHANGES
<p>Section VII. Technical Specifications</p> <p>III. Support Equipment</p> <p>1. Microbalance with RS232 Converter Max Capacity: 2.1 g/ 5.2g Readability: 0.001 mg / 0.01 mg Repeatability: (sd): 0.004 mg (for 1g) / 0.01mg Linearity: ± 0.01 mg / ± 0.02 mg Minimum weight (typical): 5mg</p>	<p>Section VII. Technical Specifications</p> <p>III. Support Equipment</p> <p>1. Microbalance with RS232 Converter Max Capacity: 2.1 g/ 5.2g Readability: 0.001 mg / 0.01 mg Repeatability: (sd): 0.004 mg (for 1g) / 0.01mg Linearity: ± 0.01 mg / ± 0.02 mg Minimum weight (typical): 5mg <u>With anti-vibration table</u></p>

Prepared by:

JENNY LOU R. FLORES
Head, BAC Secretariat

Approved by:

ATTY. RONALD E. RIMANDO
Chairperson, Bids and Awards Committee



Management
System
ISO 9001:2015
www.tuv.com
ID 9105085567

