



Republic of the Philippines  
Department of Agriculture

## SUGAR REGULATORY ADMINISTRATION

Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: [srahead@sra.gov.ph](mailto:srahead@sra.gov.ph)

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376

### SUGAR REGULATORY ADMINISTRATION

North Avenue., Diliman, Quezon City

### BIDS AND AWARDS COMMITTEE

### SUPPLEMENTAL BID BULLETIN NO. 1

This **Supplemental Bid Bulletin No. 1** dated **15 April 2025** is issued to inform the prospective bidders of the following amendments, modifications and clarifications on the items in the Philippine Bidding Documents (PBDs) for the **PROCUREMENT OF SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND NEW UV-VIS SPECTROPHOTOMETER** with Project ID No. **SRA2025-005** as previously posted in the Philippine Government Electronic Procurement System (PhilGEPS) and the Sugar Regulatory Administration (SRA) website. This Supplemental Bid Bulletin shall form an integral part of the Bidding Documents.

ORIGINAL BID DOCUMENT	CLARIFICATIONS/CORRECTIONS/CHANGES
<p><b><i>Section I. Invitation to Bid</i></b></p> <p>9. Bid opening shall be on April 28, 2025 at the given address below <b><i>SRA Social Hall, Sugar Regulatory Administration, North Avenue, Diliman, Quezon City</i></b> through Physical or Face-to-Face. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>	<p><b><i>Section I. Invitation to Bid</i></b></p> <p>9. Bid opening shall be on April 28, 2025 - <b><u>9:30 AM</u></b> at the given address below <b><i>SRA Social Hall, Sugar Regulatory Administration, North Avenue, Diliman, Quezon City</i></b> through Physical or Face-to-Face. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>
<p><b><i>Section VII. Technical Specifications</i></b></p> <p>Wavelength scanning speed : 3,000 to 2 nm/min 29,000 nm/min when survey scanning</p> <p>Lamp interchange wavelength : Automatic interchange linked to wavelength</p>	<p><b><i>Section VII. Technical Specifications</i></b></p> <p>Wavelength scanning speed : 3,000 to 2 nm/min <b><u>Between 24,000- 29,000 nm/min</u></b> when survey scanning</p> <p>Lamp interchange wavelength : Automatic interchange linked to wavelength <b><u>for halogen deuterium lamps/ no interchange needed for xenon lamps</u></b></p>

Supplemental Bid Bulletin No. 1 Supply and Delivery of One (1) Unit Brand New UV-VIS Spectrophotometer




Management System  
ISO 9001:2015  
[www.tuv.com](http://www.tuv.com)  
ID 3105085567




The interchange wavelength can be set freely in the range of 295 to 34 nm (0.1 increments).		The interchange wavelength can be set freely in the range of 295 to 34 nm (0.1 increments).	
Photometric repeatability	: Less than $\pm 0.001$ Abs at 0.5 Abs Less than $\pm 0.001$ Abs at 1 Abs Less than $\pm 0.003$ Abs at 2 Abs	Photometric repeatability	: Less than $\pm 0.001$ Abs at 0.5 Abs Less than $\pm 0.001$ Abs at 1 Abs
Light Source	: 20-W halogen and deuterium lamps Built-in light source  Auto position adjustment	Light Source	: 20-W halogen and deuterium lamps Built-in light source / <u>xenon lamp</u>  Auto position adjustment

Prepared by:

  
**JENNY LOU R. FLORES**  
Head, BAC Secretariat

Approved by:

  
**ATTY. RONALD E. RIMANDO**  
Chairperson, Bids and Awards Committee