



Republic of the Philippines  
Department of Agriculture

## SUGAR REGULATORY ADMINISTRATION

Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: [srahead@sra.gov.ph](mailto:srahead@sra.gov.ph)

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376

### SUGAR REGULATORY ADMINISTRATION

North Avenue., Diliman, Quezon City

### BIDS AND AWARDS COMMITTEE

### SUPPLEMENTAL BID BULLETIN NO. 1

This **Supplemental Bid Bulletin No. 1** dated **May 06, 2025** is issued to inform the prospective bidders of the following amendments, modifications and clarifications on the items in the Philippine Bidding Documents (PBDs) for the **Procurement of Supply and Delivery of Two (2) Units Brand-New Saccharimeter with Refractometer** as previously posted in the Philippine Government Electronic Procurement System (PhilGEPS) and the Sugar Regulatory Administration (SRA) website. This Supplemental Bid Bulletin shall form an integral part of the Bidding Documents.

ORIGINAL BID DOCUMENT			CLARIFICATIONS/CORRECTIONS /CHANGES		
<b>PROCUREMENT OF SUPPLY AND DELIVERY OF TWO (2) UNITS BRAND-NEW SACCHARIMETER WITH REFRACTOMETER</b>  <b>Section VII. Technical Specifications</b>			<b>PROCUREMENT OF SUPPLY AND DELIVERY OF TWO (2) UNITS BRAND-NEW SACCHARIMETER WITH REFRACTOMETER</b>  <b>Section VII. Technical Specifications</b>		
Technical Specifications			Technical Specifications		
Item	Specification	Statement of Compliance	Item	Specification	Statement of Compliance
		<p><small>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidder's bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</small></p>			<p><small>[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidder's bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</small></p>
2 unit	<b>PROCUREMENT OF SUPPLY AND DELIVERY OF TWO (2) UNITS BRAND-NEW SACCHARIMETER WITH REFRACTOMETER</b>  <b>Refractometer Features:</b> <ul style="list-style-type: none"> <li>Fast response time</li> </ul>		2 unit	<b>PROCUREMENT OF SUPPLY AND DELIVERY OF TWO (2) UNITS BRAND-NEW SACCHARIMETER WITH REFRACTOMETER</b>  <b>Refractometer Features:</b> <ul style="list-style-type: none"> <li>Fast response time</li> </ul>	



Management System  
ISO 9001:2015


[www.tuv.com](http://www.tuv.com)  
ID 1105085567






<ul style="list-style-type: none"> <li>• Easy to operate</li> <li>• High accuracy</li> <li>• Low sample volume required</li> <li>• No need for sample preparation</li> <li>• LED light source with 100,000 hours usage</li> <li>• Durable and chemical-resistant measuring prism</li> <li>• Built-in Peltier temperature control</li> <li>• Can be connected with saccharimeter</li> </ul> <p><b>Technical Specifications:</b></p> <p><b>Brix Scale Measurement</b></p> <p>Range : 0 to 100% Resolution : 0.01% Accuracy : 0.05%</p> <p><b>Temperature Measurement</b></p> <p>Range : 4 to 85°C Accuracy : ± 0.05°C Stability : ± 0.002°C</p> <p><b>Operational Accessories Included per Unit:</b></p> <ul style="list-style-type: none"> <li>• Digital Thermometer with flexible sensor and 0.001°C resolution</li> </ul>	<ul style="list-style-type: none"> <li>• Easy to operate</li> <li>• High accuracy</li> <li>• Low sample volume required</li> <li>• No need for sample preparation</li> <li>• LED light source with 100,000 hours usage</li> <li>• Durable and chemical-resistant measuring prism</li> <li>• Built-in Peltier temperature control</li> <li>• Can be connected with saccharimeter</li> </ul> <p><b>Technical Specifications:</b></p> <p><b>Brix Scale Measurement</b></p> <p>Range : 0 to 100% Resolution : 0.01% Accuracy : 0.05%</p> <p><b>Temperature Measurement</b></p> <p><b><u>Range</u> : <u>4 to 110°C</u></b> Accuracy : ± 0.05°C Stability : ± 0.002°C</p> <p><b>Operational Accessories Included per Unit:</b></p> <ul style="list-style-type: none"> <li>• <b><u>Digital Thermometer with flexible sensor and 0.01°C resolution</u></b></li> </ul>
--	---

Prepared by:

  
**JENNY LOU R. FLORES**  
Head, BAC Secretariat

Approved by:

  
**Atty. RONALD E. RIMANDO**  
Chairperson, Bids and Awards Committee



Management System  
ISO 9001:2015  
www.tuv.com  
ID 9105085567

