



Republic of the Philippines
Department of Agriculture

SUGAR REGULATORY ADMINISTRATION

Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: srahead@sra.gov.ph

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376

SUGAR REGULATORY ADMINISTRATION

North Avenue., Diliman, Quezon City

BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL BID BULLETIN NO. 1

This **Supplemental Bid Bulletin No. 1** dated **27 May 2025** is issued to inform the prospective bidders of the following amendments, modifications and clarifications on the items in the Philippine Bidding Documents (PBDs) for the **PROCUREMENT OF SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND-NEW LABORATORY FUME HOOD WITH ONE (1) PIECE MOLDED FIBERGLASS INTERIOR** as previously posted in the Philippine Government Electronic Procurement System (PhilGEPS) and the Sugar Regulatory Administration (SRA) website. This Supplemental Bid Bulletin shall form an integral part of the Bidding Documents.

ORIGINAL BID DOCUMENT					CLARIFICATIONS/CORRECTIONS/CHANGES				
Section I. Invitation to Bid					Section I. Invitation to Bid				
<p>2. The SRA now invites bids for the above Procurement Project. Delivery of the Goods is required by Ninety (90) Calendar Days from receipt of the Notice to Proceed (NTP). Bidders should have completed, within Five (5) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).</p>					<p>2. The SRA now invites bids for the above Procurement Project. Delivery of the Goods is required by <u>One Hundred Twenty (120) Calendar Days</u> from receipt of the Notice to Proceed (NTP). Bidders should have completed, within Five (5) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).</p>				
Section VI. Schedule of Requirements					Section VI. Schedule of Requirements				
<p>The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.</p>					<p>The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.</p>				
Item Number	Description	Quantity	Total	Delivered, Weeks/Months	Item Number	Description	Quantity	Total	Delivered, Weeks/Months
	PROCUREMENT OF SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND-NEW LABORATORY FUME HOOD WITH ONE (1) PIECE MOLDED FIBERGLASS <i>(Refer to Section VII. Technical Specifications)</i>	1		Within Ninety (90) Calendar Days upon receipt of Notice to Proceed		PROCUREMENT OF SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND-NEW LABORATORY FUME HOOD WITH ONE (1) PIECE MOLDED FIBERGLASS <i>(Refer to Section VII. Technical Specifications)</i>	1		Within <u>One Hundred Twenty (120) Calendar Days</u> upon receipt of Notice to Proceed



Note:			
Warranty: One (1) year warranty on parts and services, including Annual Preventive Maintenance and Calibration of the instrument.			
Service Support Requirements: To be serviced by factory trained service engineers with at least three (3) years experience in servicing the unit.			
Training: On-site training to be provided			

Note:			
List of installed units regardless of model within ten (10) years.			
Warranty: One (1) year warranty on parts and services, including Annual Preventive Maintenance and Calibration of the instrument.			
Service Support Requirements: To be serviced by factory trained service engineers with at least three (3) years experience in servicing the unit. Certificate of Training or Certification to be provided.			
Training: On-site training to be provided			

Section VII. Technical Specifications

Section VII. Technical Specifications

1
unit

**PROCUREMENT OF SUPPLY AND DELIVERY OF ONE
(1) UNIT BRAND-NEW LABORATORY FUME HOOD
WITH ONE (1) PIECE MOLDED FIBERGLASS INTERIOR**

Nominal Width	:	30"
Blower Requirements	:	Integral blower included
Conformance	:	ASHRAE 110, ASTM E84, CAN/CSA C22.2, CE, NFPA 45, SEFA 1, UL
Lighting	:	vapor-proof incandescent
Style	:	Benchtop
Dimensions metric	:	76.2 cm x 76.6 cm x 132.1 cm
Electrical	:	230 volts, 60 Hz

DESCRIPTION

- By-pass airflow design
- Durable glacier white powder-coated steel exterior.
- One-piece molded fiberglass interior and pre-set baffle with flame spread index less than 25 per ASTM E-84.
- 3/16" thick tempered safety glass vertical-rising sash with powder-coated aluminum handle.
- Pivoting air foil.
- Right side fixture panel factory-prepared to accept up to three service fixtures, one electrical duplex and one airflow monitor
- Built-in corrosion-resistant, direct drive exhaust blower capable of over-coming a maximum external static pressure of 0.12" H2O at 310 CFM (100fpm)
- Pre-wired 100-watt, vapor-proof incandescent lighting, light & blower switches.

1 unit

PROCUREMENT OF SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND-NEW LABORATORY FUME HOOD WITH ONE (1) PIECE MOLDED FIBERGLASS INTERIOR

Nominal Width	:	30"
Blower Requirements	:	Integral blower included
Conformance	:	ASHRAE 110, ASTM E84, CAN/CSA C22.2, CE, NFPA 45, SEFA 1, UL
Lighting	:	vapor-proof incandescent
Style	:	Benchtop
Dimensions metric	:	76.2 cm x 76.6 cm x 132.1 cm
Electrical	:	230 volts, 60 Hz

DESCRIPTION

- By-pass airflow design
- Durable glacier white powder-coated steel exterior.
- One-piece molded fiberglass interior and pre-set baffle with flame spread index less than 25 per ASTM E-84.
- 3/16" thick tempered safety glass vertical-rising sash with powder-coated aluminum handle.
- Pivoting air foil.
- Right side fixture panel factory-prepared to accept up to three service fixtures, one electrical duplex and one airflow monitor
- Built-in corrosion-resistant, direct drive exhaust blower capable of over-coming a maximum external static pressure of 0.12" H2O at 310 CFM (100fpm)
- Pre-wired 100-watt, vapor-proof incandescent lighting, light & blower switches.



<p>INCLUDES:</p> <ul style="list-style-type: none"> • Flat Solid Epoxy Work Surface 30" wide • Duplex Electrical Receptacle Kit • 30" Acid Cabinet without Shelf Fixture Kit, P-Trap Oval Cup Sink • Vent Kit for acid storage cabinet • Includes ducting • Certificate of Calibration of Fume Hood 	<p>INCLUDES:</p> <ul style="list-style-type: none"> • Flat Solid Epoxy Work Surface 30" wide • Duplex Electrical Receptacle Kit • 30" Acid Cabinet without Shelf Fixture Kit, P-Trap Oval Cup Sink • Vent Kit for acid storage cabinet • Includes ducting • <u>Certificate of Calibration of Fume Hood to be submitted after installation</u> <p>Notes:</p> <p><u>List of installed units regardless of model within ten (10) years.</u></p> <p>Warranty: <u>One (1) year warranty on parts and services, including Annual Preventive Maintenance and Calibration of the instrument.</u></p> <p>Service Support Requirements: <u>To be serviced by factory trained service engineers with at least three (3) years experience in servicing the unit. Certificate of Training or Certification to be provided.</u></p> <p>Training: <u>On-site training to be provided</u></p>
--	--

Prepared by:

MICHAEL LESTER S. CHAVEZ
Alternate Head, BAC Secretariat

Approved by:

ATTY. RONALD E. RIMANDO
Chairperson, Bids and Awards Committee

