



Republic of the Philippines
Department of Agriculture

SUGAR REGULATORY ADMINISTRATION


Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: srahead@sra.gov.ph

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376



SUPPLEMENTAL/BID BULLETIN NO. 1

Supply and Delivery of Two (2) units Refrigerated Centrifuge Bid Reference No. 2025-COB-03

With reference to the Bidding Documents for the Supply and Delivery of Two (2) units Refrigerated Centrifuge, posted at the PhilGEPS and SRA websites, this Supplemental/Bid Bulletin is being issued after having conferred with the Technical Working Group and the Provisional Member/End-user Representative, to clarify, amend, and modify some items in the provisions of the Bidding Documents, in response to the recommendations raised by prospective bidders, as follows:

Section VII. Technical Specifications			
From		To	
A.1	Rotor Capacity: 4 x 180 mL	A.1	Rotor Capacity: 4 x 180 or 250 mL
A.2	Maximum Speed: 5,500 rpm (max)	A.2	Maximum Speed: 5,500–18,000 rpm or better specification, or equivalent
A.3	Max. RCF: 5,580 x g	A.3	Max. RCF: 5,580-29,756 x g or better specification, or equivalent
A.6	Centrifugation Chamber: Corrosion-resistant, high thermal conductivity, 304 grade stainless steel	A.6	Centrifugation Chamber: Corrosion-resistant, high thermal conductivity, 304 grade stainless steel or better specification, or equivalent
A.7	Memory: 6 programs, direct access	A.7	Memory: 6 programs, direct access or better specification, or equivalent
A.10	Temperature range: -10 to +40 degree Celsius	A.10	Temperature range: -10 to +40 degree Celsius or better specification, or equivalent

Responses to Prospective Bidder's Queries:

Bidder	Item	SRA Requirement/ Specification	Recommendations	Comments
Rainphil Inc.	Section VII. Technical Specifications	Item A.1 Rotor Capacity: 4 x 180 mL	Item A.1 4 x 250 ml	GRANTED.
		Item A.2 Maximum Speed: 5,500 rpm (max)	Item A.2 Maximum speed: 18,000 rpm or better specification or equivalent specification	GRANTED.
		Item A.3 Max. RCF: 5,580 x g	Item A.3 Maximum G-force: 29,756 xg or better specification or equivalent specification	GRANTED.
		Item A.6 Centrifugation Chamber: Corrosion-resistant, high thermal conductivity, 304 grade stainless steel	Item A.6 Better specification or equivalent specification	GRANTED.
		Item A.7 Memory: 6 programs, direct access	Item A.7 Better specification or equivalent specification	GRANTED.
		Item A.10 Temperature range: -10 to +40 degree Celsius	Item A.10 Better specification or equivalent specification	GRANTED.



Management System
ISO 9001:2015
www.sra.gov.ph
02-3455-3376



		Item A.12 Drive System: Direct, brushless induction low profile motor	Item A.12 Brushless induction	DENIED. The Brushless DC motor is preferred since it has a higher efficiency and longer lifespan.
--	--	--	----------------------------------	---

The Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified, and superseded accordingly.

For the guidance and information of all concerned.

Issued this 17th day of March 2025.

ATTY. GUILLERMO C. TEJIDA III

Chairperson, Regional Bids and Awards Committee